

## VIEW LAB MONTHLY SUMMARY

Lab Name            Govt. Block PHC, Sowdapuram.O

Lab No                7443

Month                 June

Year                  2023

Constituent Group    Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	292.07 mg/dL	0.37	0.69
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.63 mg/dL	0.35	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	119.95 mg/dL	1.23	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<b>Modify Address</b>	<b>Modify Methods</b>	<b>Submit Lab Results</b>	<b>View Lab Monthly Summary</b>
<b>Yearly Summary</b>	<b>Change Password</b>	<b>Home</b>	<b>Sign-out</b>