

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Sowdapuram.O		<a href="#">Details About Robust Analysis</a>							
Lab No	7443		<a href="#">Detail About Monthly Summary</a>							
Month	May		<a href="#">Detail about SDI</a>							
Year	2023									
Constituent Group	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a> <a href="#">Print</a>							
			Date of Result Entered : 20/05/2023							
			Date of Report Published : 05/06/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	168.15 mg/dL	-1.36	0.44
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.72 mg/dL	-0.30	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	107.73 mg/dL	-1.18	0.27
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<b>Modify Address</b>	<b>Modify Methods</b>	<b>Submit Lab Results</b>	<b>View Lab Monthly Summary</b>
<b>Yearly Summary</b>	<b>Change Password</b>	<b>Home</b>	<b>Sign-out</b>