

## VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. Block PHC, Sowdapuram.O**

Lab No **7443**

Month **March**

Year **2023**

Constituent Group **Chemistry III**

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : **20/03/2023**

Date of Report Published : **04/04/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	<b>90.32</b> mg/dL	-0.17	0.31
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	<b>1.39</b> mg/dL	0.50	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	<b>104.59</b> mg/dL	-1.09	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<b>Modify Address</b>	<b>Modify Methods</b>	<b>Submit Lab Results</b>	<b>View Lab Monthly Summary</b>
<b>Yearly Summary</b>	<b>Change Password</b>	<b>Home</b>	<b>Sign-out</b>