

CALIBRATION CERTIFICATE

DATE: 04.10.2023

NAME OF CUSTOMER:

**BLOCK MEDICAL OFFICER
UPGRADED PRIMARY HEALTH CENTER
O.SOWTHAPURAM PHC,
NAMAKKAL- DT**

DESCRIPTION OF THE EQUIPMENT: CENTRIFUGE

MODEL NO: BD-53/76684

SL.NO: 76684

COMPANY: LABOTECH

INSTRUMENT USED FOR CALIBRATION: DIGITAL TACHOMETER

MODEL NO: KM-22234BL

SL.NO: S1049947

CERTIFIED BY: MSIR INDIA PVT.

REGULATOR STEEPS	INSTRUMENT SPEED	TACHOMETER READING	ALLOWABLE ERROR	ERROR OBSERVED ±	RESULT
1.	700	718	±50	18	OK
3.	2100	2075	±50	25	OK
5.	3500	3470	±50	30	OK

NEXT CALIBRATION DUE ON: 03.10.2024

TESTED BY: G DINESH

APPROVED BY: Q.A

NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.