

CUSTOMER ADDRESS: BLOCK MEDICAL OFFICER	
UPGRADED PRIMARY HEALTH CENTER	
O.SOWTHAPURAM PHC	
NAMAKKAL DT	
	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT: MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : NBS210847251	DATE OF ISSUE : 04.10.2023
RECEIPT NO : 10136	TEMPERATURE : 24.5⁰C
RECEIPT DATE : 04.10.2023	HUMIDITY : 53%
JOB CARD NO : TPM000	AIR PRESURE : 959hpa
JOB CARD DATE : 04.10.2023	MEDIUM OF WEIGHING : GRADE III WATER
	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 100-1000µl	DENSITY CORRECTION
	FACTOR[Z]: : 1.0037 µl/mg
TIPS USED : 100-1000µl	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE ACCREDITATION: VOLUME-MICROPIPETTE (100µl-1000µl)	

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT : THERMO SCIENTIFIC

SL.NO : RW02462

RESOLUTION : 0.00001g up to 82g &above is 0.0001 g

CALIBRATED BY : NEWTONBIOMEDICAL SERVICE

CALIBRATION VALIDITY : 03.10.2024

NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate3) The calibration results reported are valid at the time of and under the stated conditions of the measurements4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.

