

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE ARIYAPURAM, TENKASI-627808.		CERTIFICATE NO		TICS/2023/AP/TEN/02	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE		THERMO SCIENTIFIC	
		MODEL		10-100 µL	
CALIBRATION DATE	20/09/2023	SERIAL NO		RW09698	
CALIBRATION DUE DATE	19/09/2024	INSTALLED DEPARTMENT		LAB	
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		ID. NO		CERTIFICATE NUMBER	
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01		TSC/23-24/5424-1	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	50	49.91	± 1	PASS
		100	99.56		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					