


CALIBRATION / TESTING REPORT					
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE ARIYAPURAM, TENKASI-627808.		CERTIFICATE NO		TICS/2023/AP/TEN/03	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE		THERMO SCIENTIFIC	
		MODEL		100-1000 µL	
CALIBRATION DATE	20/09/2023	SERIAL NO		RW12106	
CALIBRATION DUE DATE	19/09/2024	INSTALLED DEPARTMENT		LAB	
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		ID. NO	CERTIFICATE NUMBER		VALIDITY
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01	TSC/23-24/5424-1		27-JUNE-2024
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	99.87	± 1	PASS
		1000	999.14		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 				Authorized Sign With Seal 