




CALIBRATION / TESTING REPORT					
<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE ARIYAPURAM, Tenkasi-627808.		CERTIFICATE NO		TICS/2023/AP/TEN/01	
		EQUIPMENT NAME		CENTRIFUGE	
		MANUFACTURE		BD INSTRUMENTATION	
		MODEL		BDI 152	
CALIBRATION DATE	20/09/2023	ASSET NO		AP-PH-001	
CALIBRATION DUE DATE	19/09/2024	INSTALLED DEPARTMENT		LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
<b>TEST EQUIPMENT</b>		<b>MANUFACTURE</b>		<b>MODEL</b>	
DIGITAL TACHOMETER		KUSAM MECO		KM2234 BL	
ELECTRICAL SAFTY ANALYZER		DATREND		V Pad-mini	
				SERIAL NUMBER	
				S1049875	
				VPM 18040011	
<b>TEST REPORT</b>					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 2	2135	± 5	PASS
		Speed 4	2578		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :		Authorized Sign With Seal			
					
<div style="text-align: center;">  </div>					
*Electrical safety test result attached					