

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	<b>Govt. PHC, Marathi Palayam</b>	<b>Details About Robust Analysis</b>
<b>Lab No</b>	<b>13051</b>	<b>Detail About Monthly Summary</b>
<b>Month</b>	<b>May</b>	<b>Detail about SDI</b>
<b>Year</b>	<b>2023</b>	
<b>Constituent Group</b>	<b>Chemistry III</b>	

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/05/2023


Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	179 mg/dL	-0.20	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	109.25 mg/dL	-0.92	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
 MEDICAL OFFICER  
 GOVT PRIMARY HEALTH CENTRE  
 MARATTIPALAYAM 632 107