

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	<b>Govt. PHC, Marathi Palayam</b>	<b>Details About Robust Analysis</b>
<b>Lab No</b>	<b>13051</b>	<b>Detail About Monthly Summary</b>
<b>Month</b>	<b>April</b>	<b>Detail about SDI</b>
<b>Year</b>	<b>2023</b>	
<b>Constituent Group</b>	<b>Chemistry III</b>	

Click on the analyte to view Graphical Data

Date of Result Entered : 19/04/2023


Date of Report Published : 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	118.42 mg/dL	-0.14	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.28 mg/dL	0.69	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	113.2 mg/dL	0.27	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
**MEDICAL OFFICER**  
**GOVT PRIMARY HEALTH CENTRE**  
**MARATTIPALAYAM/ 632 107**