

SUMMARY

Govt. PHC, Marathi Palayam

13051

August

2023

up Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
					CV	SD				
III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	183 mg/dL	0.82	0.42	
NE III	Jaffes Kinetic - Alkaline Picrate	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.93 mg/dL	1.43	0.01	
ROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22	
Interpretation										
Range										
0 +1.00	Excellent.									
to ±2.00	Good.									
to ±2.99	Accept with caution. Warning Signal.									
	Unacceptable performance. Action Signal.									

[Signature]
 MEDICAL OFFICER

GOVT PRIMARY HEALTH CENTRE
 MARATTIPALAYAM 632 107