

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





13009

Lab Name Govt. PHC, Guruvaraja Pettai Lab No

Constituent Group Chemistry III Date of Result Entered: 19/08/2023

PT item Lyophilized human serum based Date of Report Published: 01/09/2023

CLNa	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI	
SI.No						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	194 mg/dL	2.05	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	124 mg/dL	2.20	0.22

SDI Range	Interpretation		
Within -1.00 to +1.00	Excellent.		
Within ±1.01 to ±2.00	Good.		
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.		
Beyond ±3.0	Unacceptable performance. Action Signal.		

LAB ADDRESS:

Govt. PHC, Guruvaraja Pettai

Guruvaraja Pettai TAMILNADU

Coordinator Contact Details:

Dr. Pamela Christudoss

Email:clingc@cmcvellore.ac.in

CMC EQAS Coordinator

Contact Number: 0416-2283102 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******** End of Report *******