

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/03866

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ARAKKONAM BLOCK, GURUVARAJAPETTAI-631101, RANIPET DISTRICT.

SRF No.	SRF/23/00143-0003
SRF Date	04-04-2023
Date of Receipt	01-04-2023
Date of Calibration	04-04-2023
Due Date for Calibration	03-04-2024
Issue Date	05-04-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	MICROLUX
Range	5-50µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	5-50µl
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.6	Humidity (%RH)	51	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SL No.	Description	ID.No. / SL No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

SL No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.87	9.86	9.87	9.87	9.88	9.87	-0.13	0.01	0.47
		9.88	9.86	9.87	9.88	9.89				
2	30	29.76	29.77	29.78	29.79	29.76	29.78	-0.22	0.01	0.47
		29.77	29.78	29.79	29.78	29.78				
3	50	49.88	49.87	49.88	49.87	49.88	49.87	-0.13	0.01	0.47
		49.87	49.86	49.86	49.87	49.88				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
P.MYLSAMY



Authorised by:

(Quality Manager/Chief Executive)
D.VETRI SELVI

SUNSHINE BIOMEDICAL SOLUTIONS



CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/23/03864

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ARAKKONAM BLOCK, GURUVARAJAPETTAI-631101, RANIPET DISTRICT.

SRF No.	SRF/23/00143-0001
SRF Date	04-04-2023
Date of Receipt	01-04-2023
Date of Calibration	04-04-2023
Due Date for Calibration	03-04-2024
Issue Date	05-04-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100 μ l	Model	FINNPIPETTE F3
Resolution	0.2 μ l	Material	PVC
Serial Number	RW09343	Operating Range	10-100 μ l
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature ($^{\circ}$ C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature ($^{\circ}$ C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in μ l

SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (\pm)
1	10.0	9.87	9.86	9.87	9.86	9.87	9.86	-0.14	0.01	0.47
		9.86	9.82	9.86	9.87	9.86				
2	50.0	49.46	49.47	49.48	49.47	49.46	49.48	-0.52	0.01	0.47
		49.47	49.48	49.48	49.49	49.49				
3	100.0	99.37	99.38	99.39	99.31	99.40	99.40	-0.60	0.04	0.47
		99.42	99.42	99.44	99.44	99.45				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor $k=2.00$.
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
P.MYILSAMY



Authorised by:

(Quality Manager/Chief Executive)
D.VETRI SELVI

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/03867

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, ARAKKONAM BLOCK, GURUVARAJAPETTAI-631101, RANIPET DISTRICT.	SRF No.	SRF/23/00143-0004
	SRF Date	04-04-2023
	Date of Receipt	01-04-2023
	Date of Calibration	04-04-2023
	Due Date for Calibration	03-04-2024
	Issue Date	05-04-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW12481	Operating Range	100-1000µl
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition
Calibration Method Used

Temperature (°C)	23.4	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP12/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.92	99.94	99.91	99.90	99.92	99.92	-0.08	0.01	0.47
		99.94	99.93	99.92	99.91	99.92				
2	500	499.04	498.99	499.05	499.02	499.04	499.01	-0.99	0.03	0.47
		498.96	498.99	499.02	499.03	498.99				
3	1000	998.99	998.89	998.90	998.91	998.91	998.93	-1.07	0.03	0.47
		998.96	998.95	998.92	998.94	998.93				


Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,


 (Calibration Engineer)
 P. MYILSAMY


Authorised by:


 (Quality Manager/Chief Executive)
 D. VETRI SELVI



CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/23/03865

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ARAKKONAM BLOCK, GURUVARAJAPETTAI-631101, RANIPET DISTRICT.

SRF No.	SRF/23/00143-0002
SRF Date	04-04-2023
Date of Receipt	01-04-2023
Date of Calibration	04-04-2023
Due Date for Calibration	03-04-2024
Issue Date	05-04-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	MICROLUX
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.6	Humidity (%RH)	51	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	100.51	100.52	100.53	100.52	100.51	100.50	0.50	0.02	0.47
		100.49	100.48	100.49	100.49	100.48				
2	500	500.40	500.42	500.43	500.44	500.46	500.43	0.43	0.02	0.47
		500.45	500.44	500.43	500.42	500.41				
3	1000	1001.24	1001.26	1001.23	1001.22	1001.22	1001.22	1.22	0.03	0.47
		1001.18	1001.17	1001.20	1001.24	1001.26				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
P.MYILSAMY



Authorised by:

(Quality Manager/Chief Executive)
D.VETRI SELVI