

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Peenjamantai

Details About Robust Analysis

Lab No 10653

Detail About Monthly Summary

Month May

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 180 mg/dL | -0.09 | 0.44 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1854 | 1.76 | 7.48 | 0.13 | 1.8 mg/dL | 0.30 | 0.01 |
| | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1798 | 114.60 | 5.08 | 5.83 | 110 mg/dL | -0.79 | 0.27 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Outside ±3.0 | Unacceptable performance. Action Signal. |

| | | | |
|---------|-----------------|--------------------|--------------------------|
| Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Summary | Change Password | Home | Sign-out |