

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Peenjamantal
 Lab No 10653
 Month August
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	189 mg/dL	1.49	0.62
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	102 mg/dL	-1.25	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out