



# G MULTI SERVICES

78, Yahooob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin:638 452.

gmultiservicesgobi@gmail.com. Cell : 96597 20654.

Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE.,BE (ICE)  
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE

MAKE : FINNPIPETTE F3

MODEL : 10-100 $\mu$ L

PROTECT CLASS /TYPE : Risk Class II

S. No. : QW10021

Location : GPHC, PEENJAMADHAI

S.No	Measurement Parameters	Status	Remarks
1	<b>Electrical Test</b> (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	-	-
2	<b>Physical Condition Test</b> (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	<b>Performance Test</b> (a) Mean 99.2 $\mu$ l (b) A -0.6 $\mu$ l (c) Acc % -0.71 (d) CV% -0.36	OK OK OK Ok	
4	<b>Calibration Test</b> min volume 10 $\mu$ l max Volume 100 $\mu$ l  (a) A+or - 1.5 , 5 $\mu$ l (b) S 1.5 $\mu$ l max Volume 5 $\mu$ l (c) Acc+ or - 1.50, 0.50 (d) CV% 0.50, 0.20	Ok Ok Ok Ok	Tested
<b>Formulated By:</b> Edition Research Laboratory		<b>Issued By:</b>	
<b>Endorsed By :</b> Research team head  Reviewed On: 07.09.23 Next Review : 08.09.24			