



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. Block PHC, Pudukottai

Lab No 7265

Constituent Group Chemistry III

Date of Result Entered : 20/03/2023

PT Item Lyophilized human serum based

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	99 mg/dL	1.10	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	110 mg/dL	-0.16	0.28
SDI Range Interpretation										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. Block PHC, Pudukottai
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT
Pudukottai
TAMILNADU606402

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Pamela Christudoss
MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
PUDUKOTTAI-606 402,
KALLAKURICHI DT

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential