

# HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),  
SIVAKASI, VIRUDHUNAGAR-626124.  
E-mail : [hopeultrasystems@gmail.com](mailto:hopeultrasystems@gmail.com)

GSTIN: 33DLMPK3057D1Z9  
Mobile: +919487402761, +919715223192.

## CALIBRATION CERTIFICATE

Requested By

**GPHC, MANAGASERI.**

### Scope:

Equipment Name : Micropipette  
Make : Dragon Lab  
Model : 10-100µl  
Sl no : YE17AA0040433

Asset no : PHC/MS-LAB/003  
Calibration Done at : customer place  
Calibration Date : 24/08/2023  
Next call due on : 23/08/2024

### TEST RESULTS Volume µl

Sl. No.	DUC reading(µl)	Standard Reading (µl)		Correction (µl)
		Trail 1	Trail 2	
01	10	10.0	10.0	+0
02	20	20.0	20.0	+0
03	50	50.0	50.0	+0
04	80	80.0	80.0	+0
05	100	100.0	100.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

1. Electronic weighing scale(EWC/SF400C)

### Note:

1. Calibration points were selected as per customer request.
2. Each reading was taken by repeatability.
3. Results reported are valid at the time of and under the stated conditions of measurements.

