

CALIBRATION REPORT

Report No: HBMSCAL/21-24		Calibration Date: 04.09.2021
		Calibration Valid: 01.01.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, VANNIKONENDAL, THIRUNELVELL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BD INSTRUMENTATION
iii	Serial No	BDI-152
iv	Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For - Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM
1	1000	1000	0
2	1200	1209	9
3	1500	1507	7
4	2000	1999	-1

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, NavalIndia Road, Peelamedu,

CBE - 04, hbmedicals@gmail.com



S.No	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT-100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non-contact) measured has been inspected in accordance with our inspection procedures to meet manufactures specification.

Remarks

This certificate may not be reproduced other than in full Except prior permission of laboratory.

Verified by :

S. Anje



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navalndla Road, Peelamedu,

CBE - 04, hbmedicalsystm@gmail.com