



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. Block PHC, Melchithamur

Lab No: 7280

Constituent Group: Chemistry III

Date of Result Entered: 19/06/2023

PT item: Lyophilized human serum based

Date of Report Published: 04/07/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	282 mg/dL	-0.31	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	27.5 mg/dL	-1.22	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.65 mg/dL	0.47	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	108.05 mg/dL	-1.12	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

MEDICAL OFFICER

Govt Primary Health Co.
Melchithamur
Villupuram District-604205

LAB ADDRESS :
Govt. Block PHC, Melchithamur
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.
Melchithamur
TAMILNADU604203

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102