

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Government Primary Health Centre, Sikkal, Ramanathapuram Dist.	Certificate No:	AMS/LAB/289/4355
		Issuing Date:	18/08/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	SPHC/LAB/VMP/003
Make:	Finn Pipette ' F3	Department:	Laboratory
Model:	Thermo	Calibrated On:	18/08/2023
Sl. No:	QW14102	Calibration Due Suggested:	17/08/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 to 1000µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100	800.0	800.0	+00.0
2	200	200.0	200.0	+00.0
3	500	520.0	520.0	+20.0
4	800	830.0	830.0	+30.0
5	1000	1020.0	1020.0	+20.0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

amsm@gmail.com



amshelthcaremadurai@gmail.com



Approved by:

(Technical Manager)

94434 74610, 93455 56148