



CMC EQAS does not sub contract any components  
**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name Govt. PHC, Thumpalapatti  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 12606  
Date of Result Entered : 20/03/2023  
Date of Report Published : 04/04/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1878               | 91.48  | 7.46         | 6.82 | 89.2 mg/dL | -0.33 | 0.31 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1851               | 1.33   | 9.08         | 0.12 | 1.2 mg/dL  | -1.07 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1761               | 110.93 | 5.23         | 5.80 | 114 mg/dL  | 0.53  | 0.28 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
Govt. PHC, Thumpalapatti  
GPHC THUMBALPATTY SALEM  
Thumpalapatti  
TAMILNADU636204

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*Chulabala*  
மருத்துவ அலுவலர்  
அரசு ஆரம்ப சுகாதார நிலையம்  
குடாபட்டி.  
தேலம்-636 204

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. PHC, Thumpalapatti  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 12606  
Date of Result Entered : 20/04/2023  
Date of Report Published : 06/05/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value  | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|-------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |             |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1832               | 119.45 | 6.04         | 7.21 | 122 mg/dL   | 0.35  | 0.34 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1820               | 2.16   | 8.09         | 0.18 | 2.2 mg/dL   | 0.23  | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1750               | 111.61 | 5.19         | 5.79 | 110.3 mg/dL | -0.23 | 0.28 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. PHC, Thumpalapatti  
GPHC THUMBALPATTY SALEM  
Thumpalapatti  
TAMILNADU 636204

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Dr. Pamela Christudoss*  
மருத்துவ அலுவலர்  
அரசு ஆரம்ப கனகாதர நிலையம்  
துடபட்டி.  
சேலம்-636 204.

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name Govt. PHC, Thumpalapatti  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 12606  
Date of Result Entered : 20/05/2023  
Date of Report Published : 05/06/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value  | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|-------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |             |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1834               | 180.88 | 5.19         | 9.39 | 140.5 mg/dL | -4.3  | 0.44 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1854               | 1.76   | 7.48         | 0.13 | 1.7 mg/dL   | -0.45 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1798               | 114.60 | 5.08         | 5.83 | 111 mg/dL   | -0.62 | 0.27 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
Govt. PHC, Thumpalapatti  
GPHC THUMBALPATTY SALEM  
Thumpalapatti  
TAMILNADU636204

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*Pamela Christudoss*  
மருத்துவ அலுவலர்  
அரசு ஆரம்ப க்ககாதர நிலையம்  
துடபபைட்டி.  
சேலம்-636 204

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential



CMC EQAS does not sub contract any components  
**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name Govt. PHC, Thumpalapatti  
 Constituent Group Chemistry III  
 PT Item Lyophilized human serum based

Lab No 12606  
 Date of Result Entered : 20/06/2023  
 Date of Report Published : 04/07/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |       | Your Value | SDI  | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|------|------|
|       |                |                                       |  |                    |        | CV           | SD    |            |      |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1823               | 286.63 | 5.17         | 14.82 | 290 mg/dL  | 0.23 | 0.69 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1848               | 2.57   | 6.64         | 0.17  | 2.8 mg/dL  | 1.34 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1894               | 113.73 | 4.46         | 5.07  | 118 mg/dL  | 0.84 | 0.24 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
 Govt. PHC, Thumpalapatti  
 GPHC THUMBALPATTY SALEM  
 Thumpalapatti  
 TAMILNADU636294

Coordinator Contact Details:  
 Email:clinq@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 மருத்துவ அலுவலர்  
 அரசு ஆரம்ப சுகாதார நிலையம்  
 தும்பலாபட்டி  
 சென்னை-636 294

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. PHC, Thumpalapatti  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 12866  
Date of Result Entered : 19/07/2023  
Date of Report Published : 07/08/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1876               | 109.35 | 5.55         | 6.07 | 100 mg/dL  | -1.54 | 0.28 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1850               | 5.83   | 6.00         | 0.35 | 5.3 mg/dL  | -1.51 | 0.02 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1776               | 113.00 | 4.12         | 4.66 | 115 mg/dL  | 0.43  | 0.22 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

#### LAB ADDRESS :

Govt. PHC, Thumpalapatti  
GPHC THUMBALPATTY SALEM  
Thumpalapatti  
TAMILNADU636204

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*Pamela Christudoss*  
மருத்துவ அறிவுரை  
சுக அமைச்சு  
சுக அமைச்சு  
தலைநகர்-636 204

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. PHC, Thumpalapatti  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 12606  
Date of Result Entered : 19/08/2023  
Date of Report Published : 01/09/2023


| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1841               | 175.58 | 5.13         | 9.00 | 175 mg/dL  | -0.06 | 0.42 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1852               | 1.76   | 6.77         | 0.12 | 2.4 mg/dL  | 5.33  | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1764               | 113.80 | 4.07         | 4.64 | 115 mg/dL  | 0.26  | 0.22 |

| SDI Range             | Interpretation                           |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
Govt. PHC, Thumpalapatti  
GPHC THUMBALPATTY SALEM  
Thumpalapatti  
TAMILNADU636204

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

  
மருத்துவ அலுவலர்  
ஆசை ஆரம்ப ககாதர நினையம்  
குடபம்படி.  
சேலம்-636 204

  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential