



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. PHC, Chinnathampalalayam Lab No: 12178
 Constituent Group: Chemistry III Date of Result Entered: 20/09/2023
 PT Item: Lyophilized human serum based Date of Report Published: 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	317 mg/dL	0.41	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.4 mg/dL	1.05	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	108 mg/dL	-1.30	0.20
SDI Range Interpretation										
				Within -1.00 to +1.00 Excellent.						
				Within ±1.01 to ±2.00 Good.						
				Within ±2.01 to ±2.99 Accept with caution. Warning Signal.						
				Beyond ±3.0 Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Chinnathampalalayam
 THE MEDICAL OFFICER, ANDHITYUR
 Chinnathampalalayam
 TAMILNADU

Coordinator Contact Details:
 Email: cllnqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Dr. Parvitha Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed
 Data In CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****