

Name Govt. PHC, Chinnathambipalayam

[Details About Robust Analysis](#)

No 12178

[Detail About Monthly Summary](#)

Month June

[Detail about SDI](#)

Year 2023

Instituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	299 mg/dL	0.83	0.69
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	117 mg/dL	0.64	0.24

SDI Range	Interpretation
in -1.00 to +1.00	Excellent.
in +1.01 to ±2.00	Good.