



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12769

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE(MAILAM BLOCK),  
NEDIMOZHIANUR-604305,VILLUPURAM DISTRICT.

SRF No.	SRF/23/00602-0001
SRF Date	03-10-2023
Date of Receipt	03-10-2023
Date of Calibration	03-10-2023
Due Date for Calibration	02-10-2024
Issue Date	04-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	RIGHT Q PLUS
Range	10-100µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.6	Humidity (%RH)	54	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.96	9.95	9.98	9.96	9.97	9.97	-0.03	0.01	0.47
		9.98	9.96	9.97	9.98	9.95				
2	50	49.86	49.85	49.87	49.86	49.85	49.91	-0.09	0.05	0.47
		49.95	49.96	49.97	49.95	49.95				
3	100	99.98	99.97	99.98	99.96	99.98	99.96	-0.04	0.01	0.47
		99.95	99.96	99.94	99.95	99.96				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

*M. Adhiban*  
(Calibration Engineer)  
M.ADHIBAN



Authorised by:

*S. Divahar*  
(Quality Manager/Chief Executive)  
C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/12770

Page No : 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE(MAILAM BLOCK), NEDIMOZHIANUR-604305,VILLUPURAM DISTRICT.	SRF No.	SRF/23/00602-0002
	SRF Date	03-10-2023
	Date of Receipt	03-10-2023
	Date of Calibration	03-10-2023
	Due Date for Calibration	02-10-2024
	Issue Date	04-10-2023

**Details of Unit Under Calibration**

Description	MICRO PIPETTE	Make	NA
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.5	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

**Z Factor: 1.00319**


Result of Calibration in µl										
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.94	99.96	99.94	99.92	99.93	99.95	-0.05	0.02	0.47
		99.94	99.95	99.96	99.97	99.98				
2	500	499.85	499.86	499.87	499.89	499.86	499.87	-0.13	0.02	0.47
		499.86	499.87	499.89	499.85	499.89				
3	1000	999.96	999.97	999.98	999.96	999.95	999.96	-0.04	0.01	0.47
		999.98	999.94	999.96	999.95	999.95				

**Remarks**


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade 3 as specified in ISO 3696.

Calibrated By,

  
 (Calibration Engineer)  
 M.ADHIBAN



Authorised by:

  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.