

MEP Automation

55, West Pradakshnam Road, Karur - 639 001.

Calibration Certificate for Mircroscope

Customer : MO , Valayapatti , PHC , Karur dist

Calib. Date : 04.04.2023

Due date : 03.04.2024

Model : LABOMED

Serial no : 190614768

Calibraiotn Standrads Used

Name	Model	Srno
Digital Multimeter:	MASTECI	M-830 BZ

I .Measuring Reports

Test	Limit	Result
Supply Voltage (Line -Neutral)	210 - 240 V	230 V
Supply Voltage (Line -Ground)	210 - 240 V	235 V
Supply Voltage (Neutral- Ground)	< 5V	2 V

II. All Eye Piece Test Passed : Yes

III. All Optical lens Test Passed : Yes

IV. Mirror adjustment Test Passed : Yes

V. Mechanical stage Test Passed : Yes

Tested and Calibrated by

