



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab No

13139

Date of Result Entered :

20/04/2023

Date of Report Published :

06/05/2023

Lab Name

Govt. PHC, Kattu Yedaiyur

Chemistry III

Lyophilized human serum based

Constituent Group

PT Item

S/No Analyte Method / Principle Name

S/No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	122 mg/dL	0.35	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	143	111.22	4.31	4.79	111 mg/dL	-0.05	0.80
SDI Range										
Within -1.00 to +1.00										
Within ±1.01 to ±2.00										
Within ±2.01 to ±2.99										
Beyond ±3.0										
Interpretation										
Excellent.										
Good.										
Accept with caution. Warning Signal.										
Unacceptable performance. Action Signal.										

### LAB ADDRESS :

Govt. PHC, Kattu Yedaiyur  
Laboratory/ Kattuedaiyur phc  
Kattu Yedaiyur  
TAMILNADU605751

Coordinator Contact Details:  
Email: dtingc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

**MEDICAL OFFICER**  
Govt. Primary Health Centre  
Kattuedaiyur  
Kallakurichi District.