



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name  
Govt. PHC, Kattu Yedaiyur

Lab No  
13139

Constituent Group  
Chemistry III

Date of Result Entered :  
20/05/2023

PT Item  
Lyophilized human serum based

Date of Report Published :  
05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	Participants	Your Value	SDI
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	178 mg/dL	-0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.9 mg/dL	1.06
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	154	114.84	5.08	5.83	117 mg/dL	0.37

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Kattu Yedaiyur  
Laboratory Kattuedaiyur phc  
Kattu Yedaiyur  
TAMILNADU605751

Coordinator Contact Details:  
Email:cdlnqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela*  
**MEDICAL OFFICER**  
Govt. Primary Health Centre  
Kattuedaiyur  
Kallakurichi District.