




CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. UPHC, Kasthuribai Maternity Home Lab No 10920
Constituent Group Chemistry III Date of Result Entered : 18/08/2023
PT item Lyophilized human serum based Date of Report Published : 01/09/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|---|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Any Analyser (Automation / Semi Automation) | 64 | 174.34 | 5.46 | 9.51 | 138 mg/dL | -3.82 | 2.38 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Any Analyser (Automation / Semi Automation) | 62 | 1.76 | 7.18 | 0.13 | 2.3 mg/dL | 4.15 | 0.03 |
| 3 | CHOLESTROL III | CHOD-PAP III | Any Analyser (Automation / Semi Automation) | 149 | 113.96 | 4.36 | 4.97 | 130 mg/dL | 3.23 | 0.81 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. UPHC, Kasthuribai Maternity Home
THE MEDICAL OFFICER, KASTHURI BAI MATERNITY HOSPITAL, ASATH ROAD
Karur
TAMILNADU639001

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****