



CALIBRATION REPORT

| | | |
|---|-----------------|---|
| Report No: HBMS/CAL/23-24 | | Calibration Date: 27.09.2023 |
| | | Calibration Valid: 26.03.2024 |
| 1.1. CUSTOMER DETAILS: | | |
| Name and Address of the organization | | M/s, PRIMARY HEALTH CENTRE, VINATHEERTHAPURAM, SELLAPAMPATTI, NAMAKKAL. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | | |
| EQUIPMENT NAME | | PIPETTE |
| I | Manufactured by | MICROLUX |
| II | Model | 5µl - 50µl |
| III | Serial No | PHC/VRM/PIP/001 |
| IV | Location | LABORATORY |



Calibration standards used :

| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Details of the machine: machine calibrated and checked output.

| Number of measurement | Channel I (Volume) | | |
|-----------------------|--------------------|-------|-------|
| | 05µl | 20µl | 50µl |
| 1 | 05.05 | 20.08 | 50.97 |
| 2 | 05.06 | 20.07 | 20.02 |
| 3 | 05.01 | 20.05 | 20.96 |
| 4 | 05.97 | 20.99 | 20.02 |
| 5 | 05.01 | 20.08 | 20.09 |
| 6 | 05.02 | 20.95 | 20.04 |

For Heart Beat Medical Systems,
(Authorized signatory)

ID Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsysteam@gmail.com