



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date: 27.09.2023
		Calibration valid: 26.03.2024
I.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, VINAITHEERTHAPURAM, SELLAPAMPATTI, NAMAKKAL.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	MICROLUX
ii	Model	---
iii	Value	100µl-1000µl
iv	Serial No	PHC/VRM/PIP/002


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-224Q	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel I (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,



[Signature]

(Authorized signatory)

