



CALIBRATION REPORT

| | | |
|---|-----------------|--|
| Report No: HBMS/CAL/23-24 | | Calibration Date: 27.09.2023 |
| | | Calibration Valid: 26.03.2024 |
| 1.1. CUSTOMER DETAILS: | | |
| Name and Address of the organization | | M/s, PRIMARY HEALTH CENTRE, VINÁITHIERTHAPURAM, SELLAPAMPATTI, NAMAKKAL. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | | |
| EQUIPMENT NAME | | PIPETTE |
| I | Manufactured by | TIERMOSCIENTIFIC |
| II | Model | FINNPIPETTE F3 |
| III | Value | 1000µl |
| IV | Serial No | RW02795 |

Calibration standards used :

| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Details of the machine: machine calibrated and checked output.

| Number of measurement | Channel 1 (Volume) |
|-----------------------|--------------------|
| | 1000µl |
| 1 | 1000.05 |
| 2 | 1000.06 |
| 3 | 1000.01 |
| 4 | 999.97 |
| 5 | 1000.01 |
| 6 | 1000.02 |



For Heart Beat Medical Systems,

(Authorized signatory)

[Handwritten Signature]

Reg No: TN/CBE/MD42/00016

ISTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com