



Trust & Care



# heartbeat

Medical Systems <sup>TM</sup>

## CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 27.09.2023
		Calibration Valid: 26.03.2024
<b>I.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, VINAITHEERTHAPURAM, SELLAPAMPATTI, NAMAKKAL.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOMED
ii	Model	BDI-153
iii	Serial No	76707
iv	Location	LABORATORY

## CALIBRATION WORKSHEET

Tabulation For - Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	1999	-1

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range

ID Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com