

05/10/2023, 10:11

External Quality Assurance Scheme - Print Monthly Summary

CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Lab Name: Govt. UPHC, Palayapettai Lab No: 13741
 Constituent Group: Chemistry III Date of Result Entered: 16/09/2023
 PT Item: Lysophilized human serum based Date of Report Published: 05/10/2023

Sl. No.	Analyte	Method / Reagent Name	Analyser Name	No. of Participants	CV	Participant		Your score	SD	Z
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1976	3.19.85	4.82	14.55	225 mg/dL	-1.46	0.55
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	9.19	1.1 mg/dL	-1.81	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.09	4.32	124 mg/dL	2.40	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±3.00	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. UPHC, Palayapettai
 THE MEDICAL OFFICER, PALAYAPETTAI
 Palayapettai
 TAMILNADU 627008

Coordinator Contact Details:
 Email: cmiq@cmcvellore.ac.in
 Contact Number: 9416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

Pamela Christudoss
 மருத்துவ அலுவலர்
 சென்னை சி.மீ.க. நகர சுகாதார ஆயுள் நிலையம்
 திருவள்ளூர்