



## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



FC-1024

Lab Name: Govt. UPHC, Palayapeta

Constituent Group: Chemistry III

PT Item: Lyophilized human serum based

Lab No: 10741

Date of Result Entered: 17/07/2023

Date of Report Published: 07/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-PGD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	99 mg/dL	-1.74	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	92 mg/dL	-4.51	0.22
SDI Range										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution, Warning Signal						
Beyond ±3.0				Unacceptable performance, Action Signal.						

LAB ADDRESS :  
Govt UPHC, Palayapeta  
THE MEDICAL OFFICER, PALAYAPETTA  
Palayapeta  
TAMILNADU6271008

Coordinator, Contact Details:  
Email: cdmq@cmcvellore.ac.in  
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*[Signature]*  
மருத்துவ அலுவலர்  
கொழும்புத்தலை நகர சுகாதார ஆரம்ப திணைக்களம்  
திருவெறல்வெளி