



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt UPHC, Palayapatti
Constituent Group: Chemistry III

Lab No.: 10741

Date of Result Entered: 18/05/2023

Date of Report Published: 05/06/2023

PT Item: Lyophilized human serum based

Sl. No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	CV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-PD III	Robonik (Semi automatic / Fully Automatic)	1834	180.00	5.19	9.39	150 mg/dL	-3.29	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1834	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution - Warning Signal						
Beyond ±3.0				Unacceptable performance - Action Signal						

LAB ADDRESS:
Govt UPHC, Palayapatti
THE MEDICAL OFFICER, PALAYAPATTAI
Palayapatti
TAMILNADU527008

Coordinator Contact Details:
Email: ctinge@gemcvellore.ac.in
Contact Number: 0415-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

மருத்துவ அலுவலர்
மதுராங்கல நகர் காமராசு ஆயுள் நிறுவனம்
திருவருங்கலம்