CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023

PC-1024

Lab Name	Govt. UPHC, Palayapetai	Lab No	10741
Constituent Group	Chemistry III	Date of Result Entered :	15/06/2023
PT item	Lyophilized human serum based	Date of Report Published :	04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	313 mg/dL	1.78	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.8 mg/dL	1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1604	113.73	4.46	5.07	132 mg/dL	3.60	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. UPHC, Palayapetai
THE MEDICAL OFFICER, PALAYAPETTAI
Palayapetai
TAMILNADU627008

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
******* End of Report ******

மருத்துவ அலுவலர்
பாளையப்பேட்டை நகர சுகாதார ஆரம்ப நிலையம்
திருநெல்வேலி