



### CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 12.09.2023
		Calibration valid : 11.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, GOVERNMENT PRIMARY HEALTH CENTRE , PALAYAPETTAI, THIRUNELVELI
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	-
ii	Model	R303
iii	Serial No	ZBEN-13706

### CALIBRATION WORKSHEET

LOCATION: LABORATORY

#### Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1501	1
4	2000	2002	2

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range

