

| CALIBRATION REPORT | |
|---|---|
| Report No : HBMS/CAL/23-24 | Calibration Date : 06.07.2023 |
| | Calibration valid: 05.01.2024 |
| I.1.CUSTOMER DETAILS: | |
| Name and Address of the organization | M/s, PRIMARY HEALTH CENTRE, PAVITHRAM, NAMAKKAL. |
| I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | |
| EQUIPMENT NAME | |
| i | Manufactured by |
| ii | Model |
| iii | Value |
| iv | Serial No |
| | MICRO PIPETTE |
| | MICROLUX |
| | --- |
| | 100µl-1000µl |
| | LABORATORY |

Calibration standards used :

| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Calibration test Report.

| Number of measurement | Channel 1 (Volume) | | |
|-----------------------|--------------------|--------|---------|
| | 100µl | 500µl | 1000µl |
| 1 | 100.03 | 500.06 | 1000.14 |
| 2 | 100.04 | 500.09 | 1000.12 |
| 3 | 100.05 | 500.02 | 1000.10 |
| 4 | 100.09 | 500.08 | 1000.08 |
| 5 | 100.00 | 499.94 | 1000.09 |
| 6 | 100.08 | 500.07 | 1000.12 |

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com