



### CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 06.07.2023
		Calibration Valid: 05.01.2024
<b>1.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, PAVITHRAM, NAMAKKAL.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MICROLUX
II	Model	10µl
III	Serial No	PHC/PV/MP/002
IV	Location	LABORATORY

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel I (Volume)
	10µl
1	10.05
2	10.06
3	10.01
4	9.97
5	10.01
6	10.02

For Heart Beat Medical Systems

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com