



# CHRISTIAN MEDICAL COLLEGE VELLORE

*In Service of the Nation Since 1900...*



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Narayanapuram

Lab No: 13155

Month: September

Year: 2023

Consultant Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	209 mg/dL	-0.12	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	110 mg/dL	-0.84	0.20
SDI Range										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good						
Within ±2.01 to ±2.99				Accept with caution Warning Signal						
Beyond ±3.0				Unacceptable performance Action Signal						