



MONTHLY SUMMARY

Name: Govt. UPHC, Subramaniyar Street, Saidapet [Details About Robust Analysis](#)  
 ID: 9295 [Detail About Monthly Summary](#)  
 Month: January [Detail about SDI](#)  
 Year: 2023  
 Student Group: Chemistry III

[Click on the analyte to view Graphical Data](#)
[All Analyser Result](#)
[Print](#)

Date of Result Entered : 20/01/2023  
 Date of Report Published : 09/02/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1890	119.87	6.38	7.65	103 mg/dL	-2.20	0.3
UREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1865	2.21	10.02	0.22	1.9 mg/dL	-1.40	0.0
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1794	112.14	6.03	6.76	124 mg/dL	1.75	0.3

SDI Range	Interpretation
Between -1.00 to +1.00	Excellent.
Between ±1.01 to ±2.00	Good.
Between ±2.01 to ±2.99	Accept with caution. Warning Signal.
Outside ±3.0	Unacceptable performance. Action Signal.

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out