

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kalimangalam
 Lab No 12359
 Month September
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/09/2023

Date of Report Published : 05/10/2023

Sl. No.	Analyte Name	Reagent Name	Analyzer Name	No of Participants	DV	Participants		Total Value	SDI	
						CV	SD		SDI	SDI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	322.6 mg/dL	0.78	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	112 mg/dL	-0.38	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links


 MEDICAL OFFICER
 Govt. Primary Health Centre
 Kalimangalam
 Madurai Dist - 625 020