

CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 30.09.2023
	Calibration valid : 29.03.2024
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, UPGRADED PRIMARY HEALTH CENTRE, OLD SURAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNPIPETTE F3
iii Value	10 µl -100 µl
iv Serial No	RW09072

Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023


Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory)





CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date: 30.09.2023
		Calibration valid: 29.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, UPGRADED PRIMARY HEALTH CENTRE, OLD SURAMANGALAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMOSCIENTIFIC
ii	Model	FINNPIPETTE F3
iii	Value	100µl-1000µl
iv	Serial No	RW11626

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

J. S. R.
(Authorized signatory)

