

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UGPHC, Gopalapuram
 Lab No 13143
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023


Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	180.6 mg/dL	0.56	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	120 mg/dL	1.34	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER
 GOVT PRIMARY HEALTH CENTRE
 GOPALAPURAM 626 112
 VIRUDHUNAGAR - DIST