

CALIBRATION CERTIFICATE							
CERTIFICATE NO: SBS/CL/23/12859		MEDICAL DEVICES		Page No	o:1 of 1		
Issue Date		03-10-2023					
SRF No & Date		SRF/23/00626-0001 & 05-10-2023					
Receipt Date		05-10-2023					
Calibration Date		05-10-2023					
Calibration Due		04-10-2024					
Customer Name & Address							
GOVERNMENT UPGRADED PRIMARY HEA	LTH CENTRE,						
KODIKULAM, RAMNATHAPURAMDISTRICT.							
	Details of Device U	Inder Calibrati	ion (DUC)				
Description : SEMI AUTO ANALYZER		& Model	:	ALPHA & CHEM100			
Range : MULTI	Sr. N	0	:	BCAA17062643			
Resolution : MULTI	Identi	fication No	:	NA			
DUC Condition : SATISFACTORY		ion	:	LABORATORY			
Environmental Conditions & Calibration Procedure Details							
Environmental Details Temperature:25.6°C		Relative Humidity 52% RH					
Calibration Procedure No SBS/CP/MD/20		Calibration done at ONSITE					
Reference Standards Details							
S.No Description Make/ SI No:			Certificate No		V	alidity	
1 Electrical Safety Analyser Rigel Medical		L-1059	M-230809-16-4 10		0-08-2024		

## ELECTRICAL SAFETY

RESULTS					
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±		
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % ( ±)		
	>20MΩ	92	13.92		
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)		
	<5000µAfor B,BF,CF	161	10.4		
3	Enclosure Leakage	Measured values in µA	Uncertainty in % ( ±)		
	<500µAfor B,BF,CF	237	8.5		

## REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) N.VIJAY

101	DICAL 001UTION
	ns T

Authorised Signatory

Technical Manager

(C.SHANMUGARAJ)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.