

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Doddappanaickanur  
 Lab No 7087  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	110 mg/dL	-1.31	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	108 mg/dL	-0.62	0.28

## SDI Range

## Interpretation

Within -1.00 to +1.00

Excellent.

Within  $\pm 1.01$  to  $\pm 2.00$ 

Good.

Within  $\pm 2.01$  to  $\pm 2.99$ 


Accept with caution. Warning Signal.

Beyond  $\pm 3.0$ 

Unacceptable performance. Action Signal.

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Change Password](#)[Home](#)[Sign-out](#)

  
 BLOCK MEDICAL OFFICER  
 Govt. Primary Health Centre  
 DHODDAPPANAICKANUR  
 Madurai District.