

View Lab Monthly Summary

Lab Name Govt. PHC, Å Rayappanpatti
 Lab No 12718
 Month March
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 21/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	70 mg/dL	-3.15	0.31
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	51 mg/dL	-0.65	0.98
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	43	4.73	7.17	0.34	4.1 mg/dL	-1.86	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	101 mg/dL	-1.71	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within Å±1.01 to Å±2.00	Good.
Within Å±2.01 to Å±2.99	Accept with caution. Warning Signal.
Beyond Å±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)
[Yearly Summary](#)

[Modify Methods](#)
[Change Password](#)

[Submit Lab Results](#)
[Home](#)

[View Lab Monthly Summary](#)
[Sign-out](#)