

View Lab Monthly Summary

Lab Name Govt. PHC, Å Rayappanpatti
 Lab No 12718
 Month April
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 21/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	104 mg/dL	-2.14	0.34
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	43 mg/dL	0.15	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	1.8 mg/dL	-2.06	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	38	4.58	9.59	0.44	4.2 mg/dL	-0.87	0.14
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	114 mg/dL	0.41	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within Å±1.01 to Å±2.00	Good.
Within Å±2.01 to Å±2.99	Accept with caution. Warning Signal.
Beyond Å±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)
[Yearly Summary](#)

[Modify Methods](#)
[Change Password](#)

[Submit Lab Results](#)
[Home](#)

[View Lab Monthly Summary](#)
[Sign-out](#)