

View Lab Monthly Summary

Lab Name Govt. PHC, Å Rayappanpatti
 Lab No 12718
 Month May
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	178 mg/dL	-0.31	0.44
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	24 mg/dL	-1.05	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.78 mg/dL	0.15	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	40	4.98	10.94	0.55	4.2 mg/dL	-1.43	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	114 mg/dL	-0.10	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within Å±1.01 to Å±2.00	Good.
Within Å±2.01 to Å±2.99	Accept with caution. Warning Signal.
Beyond Å±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)
[Yearly Summary](#)

[Modify Methods](#)
[Change Password](#)

[Submit Lab Results](#)
[Home](#)

[View Lab Monthly Summary](#)
[Sign-out](#)